Indian Institute of Technology

Department of Biological Sciences and Bioengineering

Kanpur 208 016

0512-259 4043 / Fax 0512-259 4010

Email: subbu@iitk.ac.in

K. Subramaniam Professor

AND THE OF THE OWNER OWNER OF THE OWNER OWNE

January 30, 2012

Closing date: February 8,2012

REF – IITK/BSBE/2011-12/PG/KS/1

Extension of the Closing date

With respect to the quotations sought on January 13, 2012, please note that the closing date has now been extended to February 8, 2012.

Following is a copy of the earlier announcement:

Request for submission of quotation for TWO liquid nitrogen-based cryopreservation tanks.

Please send a quotation in triplicate, in a sealed envelope, to the undersigned for the supply of the following:

- 1. Wide-mouthed liquid nitrogen tanks for storing 2-ml vials in 100-tube capacity square boxes.
- 2. Static holding capacity should be above 120 days.
- 3. Evaporation rate should be less than 1 litre / day.
- 4. Two tanks are required one with a capacity of 6000 vials or above and, another with a capacity of about 3000 vials.
- 5. Should include roller base and racks, but without boxes.
- 6. On-site service-based warranty for 2 years.

Your quote should mention the following:

- Point wise technical compliance statement. (Mentioning the page No of Catalogue)
- Validity at least for 90 days.
- Warranty/AMC period for the instrument.
- FOB (indicating port of shipment) and CIF (New Delhi) values separately.
- The quote should cover insurance for transport up to Kanpur.
- Indian agency commission if applicable (should be certified by the principal if no agency commission is applicable).
- Authorization certificate from the principal if you are an Indian agent.
- Terms and conditions for the payment, including the banker's name of the principal and the account number, if any, for electronic transfer.
- Include proprietary item certificate if applicable.

Sh bramanin

Please include technical literature to support your product. We look forward to your best offer for a model that meets the above mentioned criteria.

Sincerely,

K. Subramaniam